



SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 670-671-672, Fax: +91 - 2423 - 258676,
Email : hospital.purchase@sai.org.in/ hospital.store@sai.org.in

SHRI SAIBABA HOSPITAL, SHIRDI ANNEXURE "A"

Requirement of Syringe Pump for Casualty of our Shri Saibaba Hospital.

Sr. No.	Description	Qty.	Rate Per in Rs. (Incl. of all Taxes & Duties)	Total Amt. in Rs.
	<ul style="list-style-type: none">Graphic LCD display for Running Mode message display ,Syringe Type/Size, Delivery Rate, Infused Volume, Mains/Battery Operation, Real time pressure curve to detect occlusion of venous line, Remaining battery timeSyringe type- 5ml to 50ml/60ml and setting for all type syringe brandsSyringe Loading Technique: Front Bottom loading techniqueFlow Rate Range- Min 0.01ml/Hr. to 200ml/Hr. or moreAccuracy-Driver accuracy +/- 2 %Infusion Modes: Rate Mode, Time Mode, Body Weight Mode, Night ModeRapid Mode for Purge or Bolus- 1-1200 ml/Hr. (User Selectable)Volume infused Pre-selection- 1 to 999.9 ml in 0.1 ml increments.Occlusion Detection- 3 Selectable Occlusion pressure limit for different therapy up to 1.2 barAlarms -(Audible & visible alarms): Occlusion Pressure alarm , Syringe Empty & Volume infused alarm , Internal Malfunction , Battery Charge Low , Syringe Disengaged & Incorrectly place alarmClassification: Type CF – protected against defibrillation, Protection class II & (IP 22) Splash Proof and EMC EN60601-1 or betterBattery : Ni-Cd rechargeable battery with back-up time of min 10 HrsPower supply: 170-250 V, 50/60 HZWeight: less than 2 KgAdditional Features: History of infusion data, operational alarms and device alarms must be available Up gradable to Fluid management system;In-built maintenance counter Automatic bolus reduction system to avoid accidental bolus delivery after occlusion incident Vertical Stacking of 3 pumps together, for clearly structured fixation, must be possible with locking mechanism for optimum space management Option for RS-232 Connection for PDMS Ambulance power supply connectionAfter selection of parameter key pad locking available for security purpose with option to keep it off. Have facility to set time duration for enabling keypad lockingAll Technology should be mandatorily with safety approval- USFDA / UL / CE for Device Safety approval. Approval certificate must be submitted for the total unit.Warranty- 1 Years. It shuld include PMS visits-4 Nos. and unlimited Breakdown call and calibration whenever necessary.LAMC Cost- LAMC should be including taxes. It should be fixed for next 5 years.All the spare parts and consumables must be available with supplier for next 10 year from the date of installation.Cost of Consumables to be fixed for warranty period.	2		

Medical Director,
Shri Saibaba Sansthan Trust's
Hospital, Shirdi





SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 670-671-672, Fax: +91 - 2423 - 258676,
Email : hospital.purchase@sai.org.in/ hospital.store@sai.org.in

.SSS/WS/SSBH/STORES/(P/S)/ /2020
DATE:

To,

SUB: - Quotation for supply of Syringe Pump for Casualty Dept. of Shri Saibaba Hospital.

Dear Sir,

We have to purchase Syringe Pump for Casualty Dept. of our Shri Saibaba Hospital, Shirdi as per Technical Specification given in enclosed Annexure "A" & the following terms and conditions.

Terms & Condition

- 01) The Syringe Pump will have to be supplied F.O.R. Shirdi.
- 02) The rate should be inclusive of GST, transport charges, T.O.T. Etc.
- 03) You will have to mention name of the company, packing, rate per unit, total cost etc.
- 04) Packing, forwarding, and other charges will not be paid by the Sansthan.
- 05) The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m. exopt Sunday.
- 06) Please note that if taxes are not mentioned then rates quoted will be considered as including all taxes & F.O.R. Shirdi.
- 07) You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
- 08) In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
- 09) Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
- 10) Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.
- 11) The warranty should be 01 year.
- 12) In case supply is not made within specified time, Sansthan will be free to purchase the material from open market. In such case the supplier will have to bear the cost difference and in this case Sansthan will not bound to provide the deduction details and supplier will be blacklisted for minimum 3 years from Shri Saibaba Sansthan Trust without any intimation or prior notice or supplier will suffer from following penalty clause also-

Category of Stores	Penalty Amount
The case of an order not exceeding Rs. 2.00 Lakh in value	At the rate of ½ % per week subject to maximum limit of 10%
In case of an order of Rs.2.00 Lakh and above	At the rate of ½% per week subject to maximum limit of 5%

Page No. 2/-





SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba•hospital@sai•org•in

Page No.2

- 13) The product should be time tested, branded and should be OEM.
- 14) The quotation must be clearly mention on the sealed envelops "Syringe Pump" for Shri Saibaba Hospital.
- 15) The quotation in sealed cover may please be submitted to our Shirdi office on or before Dt. 28/03/2020.
- 16) In case of any dispute the same will be subject to the jurisdiction of the competent court at Rahata / Kopargaon.

Thanking you and with blessings of Shri Saibaba.

Yours sincerely,

Medical Director,
Shri Saibaba Sansthan Trust
Hospital, Shirdi

