



PTR NO. U-69, AHMED NAGAR

SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2123 - 258 671-672-673, Fax: +91 - 2123 - 258660, Telegram: Sainath, Shirdi.
Email: saibaba-hospital@sat-org.in

NO.SSS/WS/SSBI/STORES/(P/S)/ **111** /2025DATE:
12 9 APR 2025

To,

SUB: - Quotation for Purchase of Portable X-ray Machine for Dental Department of Shri Saibaba Hospital, Shirdi.

Dear Sir,

We have to purchase following Portable X-ray Machine for Shri Saibaba hospital, Shirdi as per Technical Specification & the terms and conditions.

Sr. No.	Description	Qty.	Rate Per in Rs.(inc.GST)	Total Amt. in Rs. (inc.GST)
1	Portable X-ray Machine	01		
	Preferred Make:- Vatech, Genoray, Carestream (Kodak)			
	Dental X-ray unit (65 KV, 2.5 mA) hand held with RVG (Sensor with software) should be supplied with adequate and compatible computer system with latest operating system i.e. desktop of latest version and with suitable printer & UPS.			
	Dental X-ray Machine:			
	1. Should be portable model that can be used as hand held unit.			
	2. Should have a X-ray tube current of minimum 2.5 mA and 65 KV fixed.			
	3. Should have a constant potential minimum 20 Khz high frequency X-ray generator.			
	4. Should have an exposure timer of minimum 0.05 to 1 seconds			
	5. Focal spot size should be 0.4mm.			
	6. Should have total filtration min. 1.5 mm Al			
	7. Should be compatible for digital radiograph.			

2023-24 W-Quotation Letter



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014

Telephone: +91 - 22 - 24166556, Fax: +91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.

Website: <http://www.shrisaibabasansthan.org>, E_mail: saibaba_anr@sancharnetin/saisandesh_anr@sancharnetin



PTR. NO. F 69, AHMED NAGAR

SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba-hospital@ssai.org.in

Sr. No.	Description	Qty.	Rate Per in Rs.(inc.GST)	Total Amt. in Rs. (inc.GST)
8.	Should operate on DC supply & SMF Battery should be replaceable			
9.	Should have source to skin distance about 200mm			
10.	Should be provided with Back scatter Shield and Carry case			
11.	The quoted model and tube should be AERB type approved and relevant copies of the certificate should be attached with the bid.			
12.	Should have safety certificate from a competent authority European CE / FDA (US) . Copy of the certificate / test report shall be produced along with the technical bid.			
13.	Machine should be Approved by AERB.			
	RVG Specifications:			
1.	CCD / Advanced / Super CMOS technology. Protected optical fiber.			
2.	Sensor size : Size No. 1 (Universal) which should be convenient and comfortable for Adult as well as Pediatric.			
3.	No. of Pixels minimum 20 lacs /mm (2 million) [true resolution]			
4.	Pixel size in the range of 18.5 X 18.5 micron To 20 X 20 micron.			
5.	High quality clear image			
6.	Sensor should be Slim & with rounded corner			
7.	Sensor should have 100 % active area			
8.	Sensor Exposure life should be minimum 4 lakhs exposure.			
9.	Should provide User friendly software TWAIN compatible software such as IOC			
10.	Sensor cable length should be 3 meters and reinforced for durability & reliability. [Fiber optic & scintillation tech.]			

2023-24 W-Quotation Letter



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014
Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798, Telegram: Sainiketan, Mumbai.
Website: <http://www.shrisaibabasansthan.org>, E-mail: saibaba_anr@sancharnetin/saisandesh_anr@sancharnetin



PTR NO : E-69, AHMED NAGAR

SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
E-mail : saibaba-hospital@zsai-org.in

Sl. No.	Description	Qty.	Rate Per in Rs.(inc.GST)	Total Amt. in Rs. (inc.GST)
11.	USB standard: USB 2.0 high speed			
12.	Number of cable bends (+/- 90 degree) max 1,00,000 times			
	Image Processing Functions:			
1.	Zoom glass, Zoom & Pan, Horizontal & Vertical Flip			
2.	Image Enhancement & Ultra Enhancement (Histogram Equalization)			
3.	Measurement of Line, Point & Curve			
4.	Image Invert, Rotation & Comments			
5.	3 D Colour Rendering & Bone Densitometry Indication			
6.	ROI Study & Image Fusion			
7.	Digital Diagnostic image quality			
8.	DICOM Compliant-Easy transfer of images & remote Analysis			
9.	Patient records management for easy clinical decision			
10.	Report Generation & Print capability			
11.	Images & diagnostics can be shared with patient on a CD with pre-loaded imaging software			
	Computer system specification:			
1.	CPU: (min.) 1Ghz,32 or 64 bit, Intel Dual core or Latest			
2.	Operating system : Windows 7 or latest			
3.	RAM: min. 2 GB			
4.	Hard Disk : min. 500 GB			
5.	USB Ports: USB 2.0 HS-2 Nos.			
6.	CD/DVD Drive			
7.	Monitor: 17" LED			
8.	Video display Adapter: 4 MB RAM, 1024 x 800 Display			
9.	Keyboard & Mouse : standard			
10.	UPS: 1 KVA with 30 min. Back up			

2023-24 W-Quotation Letter



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014
Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.
Website: <http://www.shrisaibabasansthan.org>, E-mail: saibaba_anr@sancharnet.in/saisandesh_anr@sancharnet.in



SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI Post, Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba.hospital@saib.org.in

Sr. No.	Description	Qty.	Rate Per in Rs.(inc.GST)	Total Amt. in Rs. (inc.GST)
11.	Protective sheaths should be provided.			
12.	All Technology should be mandatorily with safety approval- USFDA / UL & EuropeanCE for Device			
13.	Safety approval. Approval certificate must be submitted for the total unit.			
14.	The product should be time tested, branded and should be OEM			
15.	The life of the equipment should be mentioned			
16.	Warranty: 5 Years			
17.	CMC Cost- CMC cost should be including Taxes. It should be fixed for next 5 years.			
18.	Cost of Consumables to be fixed for warranty period.			

Terms & Conditions

- 01) The Portable X-ray Machine will have to be supplied F.O.R. Shirdi.
- 02) The rate should be inclusive of GST, transport charges, T.O.T. etc.
- 03) You will have to mention name of the company, packing, rate per unit, total cost etc.
- 04) Packing, forwarding, and other charges will not be paid by the Sansthan.
- 05) The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m. except Sunday.
- 06) Please note that if taxes are not mentioned then rates quoted will be considered as including all taxes & F.O.R. Shirdi.
- 07) You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
- 08) In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
- 09) Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
- 10) Any accessory / accessories not mentioned specifically but required to make system in working condition will have to be supplied without any extra cost.
- 11) Minimum warranty should be 05 year.
- 12) Payment will be made on receipt of Order Acceptance, GST no, Original invoice, etc.

2023-24 W-Quotation Letter



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014
Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.
Website: <http://www.shrisaibabasansthan.org> E-mail: saiibaba_anr@sancharnetin/saisandesha_anr@sancharnetin



SHRI SAIBABA SANSTHAN TRUST, SHIRDI

PER. NO. 1/09, AHMEDNAGAR

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba-hospital@sanl.org.in

- 13) In case supply is not made within specified time, Sansthan will be free to purchase the material from open market. In such case the supplier will have to bear the cost difference and in this case Sansthan will not bound to provide the deduction details and supplier will be blacklisted for minimum 3 years from Shri Saibaba Sansthan Trust without any intimation or prior notice or supplier will suffer from following penalty clause also-

Category of Stores	Penalty Amount
The case of an order not exceeding Rs. 2.00 Lakh in value	At the rate of ½ % per week subject to maximum limit of 10%
In case of an order of Rs.2.00 Lakh and above	At the rate of ½% per week subject to maximum limit of 5%

- 14) The product should be time tested, branded and should be brand new.
15) Those order will be finalized for purchase will have to provide one piece as sample of each item for approval.
16) The quotation must be clearly mention on the sealed envelops " Portable X-ray Machine" for Shri Saibaba Hospital.
17) The quotation in sealed envelope may please be submitted to our Shirdi office on or before Dt. 12/05/2025.
18) In case of any dispute the same will be subject to the jurisdiction of the competent court at Rahata / Kopargaon.

Thanking you and with blessings of Shri Saibaba.

Yours sincerely,

Medical Director,
Shri Saibaba Sansthan Trust
Hospital, Shirdi

