



# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba•hospital@sai•org•in

NO.SSS/WS/SSBH/STORES/(P/S)/ /2018  
Date:

To,

-----  
-----  
-----  
-----

**Sub : - Quotation for Purchase of 10 KVA UPS for ICCU Department  
of Shri Sainath Hospital, Shirdi.**


Dear Sir,

We have to purchase 10 KVA UPS for ICCU Dept. of Shri Sainath Hospital, Shirdi given in enclosed Annexure "A" & the following terms and conditions.

**Terms and Condition**

1. The 10 KVA UPS will have to be supplied F.O.R. Shirdi.
2. The rate should be inclusive of GST, transport charges, T.O.T. Etc.
3. Please note that if taxes are not mentioned then rates quoted will be considered as including all taxes & F.O.R. Shirdi & If we demand the rates with AMC /CMC & supplier fails to give rates with AMC / CMC then concern quotation may liable for rejection.
4. You will have to mention name of the company, packing, rate per unit, total cost etc.
5. Packing, forwarding, and other charges will not be paid by the Sansthan.
6. The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
7. In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
8. You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
9. In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
10. Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
11. All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the wards nor after the award for any part or accessory.

Page No. 2/-






# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 02423 - 258 671-672-673, Fax: +91 -02423 - 258676, Telegram: Sainath, Shirdi.  
Email : [hospital.store@sai.org.in](mailto:hospital.store@sai.org.in), [hospital.purchase@sai.org.in](mailto:hospital.purchase@sai.org.in)

Page No. 2

12. Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.
13. The supplier has to mention the one time visit charges for breakdown and servicing.
14. Equipment mentioned in the Technical specification shall operate on 230V  $\pm$ 10%, 50Hz, single phase electric supply if not mentioned specifically.
15. The necessary protective relaying / circuitry shall be there with the machines.
16. The equipment shall have valid CE marking and US FDA approval (Wherever necessary) and documentary evidence to that effect shall be submitted.
17. The equipment shall be having warranty of 01 year. The warranty and CMC shall cover the batteries and lamps irrespective of whether those are treated as consumables or otherwise.
18. Minimum 4 Nos. of preventive maintenance visits and unlimited breakdown call visits as per requirement must be attended in the warranty and AMC/CMC period. The services to the equipment in warranty and CMC period will be considered as same.
19. The equipment should be provided with one hard copy in original of the detailed service manual and operation manual. Further, a soft copy is also required.
20. Repairs and maintenance training to two Biomedical Engineers should be given for one week or more as per requirement.
21. The equipment must be tropicalized as below-
  - Working Temperature : Max.45<sup>0</sup>
  - Storage Temperature : Max.50<sup>0</sup>C,
  - Relative humidity for working : 40 - 60%,
  - Relative humidity for Storage : Max 90%
22. Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to SSST officials as mentioned elsewhere in the tender specifications.
23. The supplier has to submit user list with contact telephone numbers for the model quoted.
24. The supplier has to give past performance certificate of the quoted equipment.

Page No. 3/-





## **SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

**SHIRDI:** Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
**Telephone:** +91 - 02423 - 258 671-672-673, **Fax:** +91 -02423 - 258676, **Telegram:** Sainath, Shirdi.  
**Email :** [hospital.store@sai.org.in](mailto:hospital.store@sai.org.in), [hospital.purchase@sai.org.in](mailto:hospital.purchase@sai.org.in)

Page No. 3

25. The supplier should submit all the Datasheets of the quoted model for technical comparison.
26. The product should be time tested, branded and should be OEM.
27. The life of the equipment should be mentioned.
28. The Supplier must quote details of Technical Specification along with coloured brochure. In absence of these details, quotation will not be considered.
29. The quotation must be clearly mention on the sealed envelope **“Quotation for 10 KVA UPS for ICCU Department of Shri Sainath Hospital”**.
30. The quotation in sealed cover may please be submitted to our Shirdi office on or before **Dt. 20/06/2018**.

Thanking you and with blessings of Shri Saibaba.

Yours faithfully,

Medical Director,  
Shri Saibaba Hospitals, Shirdi





# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 02423 - 258 671-672-673, Fax: +91 -02423 - 258676, Telegram: Sainath, Shirdi.  
Email : [hospital.store@sai.org.in](mailto:hospital.store@sai.org.in), [hospital.purchase@sai.org.in](mailto:hospital.purchase@sai.org.in)

## Shri Saibaba Hospital, Shirdi

Technical Specification for 10 KVA UPS for ICCU Department of  
Shri Sainath Hospital  
(Preferred Make :- Emerson, APC, Socomec, Luminous, Cosole or Equivalent )

### Annexure "A"

Sr. No.	Description	Qty.	Rate Per Incl. of All Tax	Total Amt. in Rs.
1	<b>10 KVA UPS - Single Phase</b>	01		
	• IGBT based converter and Inverter			
	• 100 % isolation between input and output			
	• High AC to AC efficiency			
	• Power factor >0.98			
	• Battery : 12Volt/18 AH, SMF, ISI marked, Stand to be provided for battery			
	• Battery charging should be possible even at low voltage up to 180 Volt			
	• Back up: Minimum 1 Hr.			
	• Input voltage: 170 Volt to 270 Volt, Single phase			
	• Input Frequency:45 Hz to 60 Hz			
	• Output : 230 V, 50Hz, Single Phase, Pure Sine wave, Stabilized with spike Guard Surge Protection			
	• Output sockets: 230 V, 15 Amp. = 02 Nos.			
	• Display with indicator for AC mode, Battery mode, Battery charging, low battery, Bypass, Overload, Fault, Input & output voltage, input & output frequency, % Load on UPS.			
	• Output Voltage Regulation max. $\pm 1\%$			
	• Output Frequency regulation : $\pm 0.1$ Hz			
	• Transfer Time: 0 msec.			





# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 02423 - 258 671-672-673, Fax: +91 -02423 - 258676, Telegram: Sainath, Shirdi.  
Email : [hospital.store@sai.org.in](mailto:hospital.store@sai.org.in), [hospital.purchase@sai.org.in](mailto:hospital.purchase@sai.org.in)

Page No. 2

Sr. No.	Description	Qty.	Rate Per Incl. of All Tax	Total Amt. in Rs.
	<ul style="list-style-type: none"><li>Overload Capacity : 125% for 15 min. 150 % for 60 Seconds</li></ul>			
	<ul style="list-style-type: none"><li>Protection: Overload, Battery Overcharge, Battery Discharge, Output low, Short Circuit, Over Temperature, Surge</li></ul>			
	<ul style="list-style-type: none"><li>Noise Level: &lt; 50 dB</li></ul>			
	<ul style="list-style-type: none"><li>Output must be provided through Inverter battery</li></ul>			
	<ul style="list-style-type: none"><li>All Technology should be mandatorily with safety approval - CE, IEC. Approval certificate is to be submitted in addition to CE Certificate for the total unit.</li></ul>			
	<ul style="list-style-type: none"><li>The product should be time tested, branded and should be OEM</li></ul>			
	<ul style="list-style-type: none"><li><b>Warranty</b> - 1 Year</li></ul>			
	<ul style="list-style-type: none"><li><b>AMC/CMC Cost-</b> AMC / CMC cost should be including Taxes. It should be fixed for next 5 years.</li></ul>			
	<ul style="list-style-type: none"><li>Cost of Consumables to be fixed for warranty period.</li></ul>			
	<ul style="list-style-type: none"><li>Warranty &amp; CMC should include PMS visits - 4 Nos. and unlimited Breakdown calls with spare parts &amp; calibration whenever necessary free of cost.</li></ul>			

Medical Director,  
Shri Saibaba Hospitals, Shirdi

