

**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).

Telephone: +91 - 2423 - 258 671 / 676.

E-mail : hospital.purchase@sai.org.in

NO.SSS/WS/SSBH/STORES/(P/S)/ 48 /2023

DATE:

= 8 APR 2023

To,

Subject:- Quotation for Purchase of Deep Freezer -40 °C for Shri Sainath Blood Bank.

Dear Sir,

We have to purchase Deep Freezer -40 °C for Shri Sainath Blood Bank, Shirdi as per Technical Specification given in enclosed Annexure "A" & the following terms and conditions.

Terms and Condition

01. The Deep Freezer -40 °C will have to be supplied F.O.R. Shirdi.
02. The rate should be inclusive of GST, transport charges, T.O.T. etc.
03. You will have to mention name of the company, packing, rate per unit, total cost etc.
04. Packing, forwarding, and other charges will not be paid by the Sansthan.
05. The delivery of the ordered material will have to be made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
06. In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
07. You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
08. In case of breakage/ leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
09. Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any as specified in the quotation & after satisfactory installation & working it & Subjectmission of order Acceptance, Security deposit, Agreement etc. only. No advance payment will be made against the supply.
10. The equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import Subjectstitution is permitted neither before the wards nor after the award for any part or accessory.

D Drive 2021-G-Quotation-Letter-Head-P.No.4



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014

Telephone: +91 - 22 - 24166556, Fax: +91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.

Website: <http://www.shrisaibabasansthan.org>, E_mail: sarbaba_anr@sancharnet.in/saisandesh_anr@sancharnet.in



SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).

Telephone: +91 - 2423 - 258 671 / 676,

E-mail : hospital.purchase@sai.org.in

...02...

11. Any accessory/ accessories not mentioned specifically but required to put Machine in working condition will have to be supplied without any extra cost.
12. The supplier has to mention the one time visit charges for breakdown and servicing.
13. Equipment mentioned in the Technical specification shall operate on 230V \pm 10%, 50Hz, single phase electric supply if not mentioned specifically. The necessary protective relaying/ circuitry shall be there with the machines.
14. The equipment shall have valid CE marking and US FDA approval (Wherever necessary) and documentary evidence to that effect shall be Subjectmitted.
15. The equipment shall be having warranty of one year. The warranty and LAMC shall cover the batteries and lamps irrespective of whether those are treated as consumables or otherwise.
16. You will have given LAMC Charges 0.5% & CMC Charges 5% per year after order Valve (Excluding Taxes).
17. Minimum 4 Nos. of preventive maintenance visits and unlimited breakdown call visits as per requirement must be attended in the warranty and LAMC period.
18. The services to the equipment in warranty and LAMC period will be considered as same. The warranty will start from the date of actual satisfactory working of Machine & thereafter LAMC/ CMC will due for further perod as per P.O. or Agreement.
19. The equipment should be provided with one hard copy in original of the detailed service manual and operation manual. Further, a soft copy is also required.
20. Repairs and maintenance training to two Biomedical Engineers should be given for one week or more as per requirement.
21. The equipment must be tropicalized as below-
Working Temperature : Max.45^o
Storage Temperature : Max.50^oC,
Relative humidity for working : 40 - 60%,
Relative humidity for Storage : Max 90%
22. Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to Sansthan officials as mentioned elsewhere in the tender specifications.
23. The supplier has to Subjectmit user list with contact telephone numbers for the model quoted.
24. The supplier has to give past performance certificate of the quoted equipment.





PTR. NO. : E-69, AHMED NAGAR

SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 02423 - 258 671-672-673, Fax: +91 -02423 - 258676, Telegram: Sainath, Shirdi.
Email : hospital.purchase@sai.org.in

...03...

25. The supplier should submit all the Datasheets of the quoted model for technical comparison.
26. The product should be time tested, branded and should be OEM.
27. The life of the equipment should be mentioned.
28. The Supplier must quote details of Technical Specification along with coloured brochure. In absence of these details, quotation will not be considered.
29. The quotation must be clearly mention on the sealed envelope "Quotation for Deep Freezer -40 °C for Shri Sainath Blood Bank".
30. The quotation in sealed cover may please be submitted to our Shirdi office on or before Dt. 25/4 /2023.

Thanking you and with blessings of Shri Saibaba.

Yours faithfully,

Medical Director,
Shri Saibaba Hospital, Shirdi



**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 02423 - 258 671-672-673, Fax: +91 - 02423 - 258676, Telegram: Sainath, Shirdi.
Email : hospital.purchase@sai.org.in

Annexure "A"**Requirement of Deep Freezer - 40°C for Shri Sainath Blood Bank.**

Sr. No.	Technical Description	Qty.	We Comply	Rate Per in Rs. (Incl. all Taxes)	Total Amt.in Rs.
	Deep Freezer - 40 ° C	1	Make - Model -		
1	Should have a Chamber temperature range between -20 ° C to -40° C				
2	Should have an input voltage of 220V/50Hz, Single phase Ac Should have an inner chamber storage volume of 400L with a total storage capacity of 400 no's of 250ml plasma bags				
3	Should comply IEC safety/EMI/EMC standards, ISO 9001, ISO 13485 certified and CE marked				
4	Should include caster wheels as a standard feature Must have six inch, 7-day, ink-less pressure sensitive circular chart recorder and accuracy should be +/- 1 degree Celsius Must have a Heavy gauge , cold rolled annealed exterior with powder coating Should have adjustable stainless steel trays to place the bags.				
5	Should have stainless steel interior doors with magnetic latches that ensure secure storage and less cold air loss during opening and closing of doors. Refrigerating fluid & PUF insulation should be CFC free The door insulation should have a PUF thickness of 125 mm with rubber gasket sealing to ensure more temperature holding time and reduced power consumption.				

D Drive: 2021-G-Quotation-Letter-Head-P.No 10



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014

Telephone: +91 - 22 - 24166556, Fax: +91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.

Website: <http://www.shrisaibabasansthan.org> E_mail: saibaba_anr@sancharnetin/saisandesh_anr@sancharnetin



PTR. NO. : E-69, AHMED NAGAR

SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 02423 - 258 671-672-673, Fax: +91 -02423 - 258676, Telegram: Sainath, Shirdi.
Email : hospital.purchase@sai.org.in

Sr. No.	Technical Description	Qty.	We Comply	Rate Per in Rs. (Incl. all Taxes)	Total Amt.in Rs.
6	Should have hotline around the mouth of cabinet to prevent moisture condensation should have 24 hours Battery backup for TRCU that keeps the temperature monitoring uninterrupted.				
7	Should have user-friendly alarms and indications to display all possible abnormalities should have display of 4*7 Segment LED with 0.5°C display resolution.				
8	Should have a power fail alarm as a standard feature				
9	Should have audible and visual high and low temperature alarms as a standard feature				
10	Should have a chart recorder with chart duration of 7 days and chart resolution of 3° C Must have a High capacity air cooled condenser Servo controlled voltage stabilizer of 4 KVA				
11	Should be included should have hermetically sealed compressor which automatically regulate according to the freezer should have pressure sensitive mechanism to protect the compressor in long run.				
12	Company should be ISO 9001:2008 & ISO 13485:2003 certified company for manufacturing and supply of Electro-Medical Equipment. 24. Unit should be CE certified by European Notified body.				
13	Unit should be provided with standard accessories.				

D: Drive 2021-G-Quotation-Letter-Head-P No 11



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014
Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.
Website: <http://www.shrisaibabasansthan.org>. E_mail: sai_baba_anr@sancharnet.in/saisandesh_anr@sancharnet.in



PTR. NO. - E 69, AHMED NAGAR

SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 02423 - 258 671-672-673, Fax: +91 -02423 - 258676, Telegram: Sainath, Shirdi.
Email : hospital.purchase@sai.org.in

Sr. No.	Technical Description	Qty.	We Comply	Rate Per in Rs. (Incl. all Taxes)	Total Amt.in Rs.
14	The company supplying the instrument should have a good track record.				
15	Excellent service and distributor network all over India.				
16	All technology should be mandatorily with safety approval-USFDA/ UL. Approval certificate is to be submitted in addition to CE Certificate for the total unit (USFDA & European CE Certificate).				
17	Warranty- 2 Years warranty. It should include PMS Visits-4 Nos. Breakdown call & Calibration whenever necessary free of cost.				
18	LAMC Cost-LAMC Cost should be including Taxes. It should be fixed for next 5 years.				
19	List of consumables must be supplied along with. Cost of consumables to be fixed for warranty period				
20	Buyback Cost of old Deep Freezer -40 °C - Qty. 01 No. Make - Remi Model - RPFV - 240				

Note :- Please tick mark if you are complying our specification, if there is any deviation then Write in each row.

Medical Director,
Shri Saibaba Sansthan Trust
Hospital's, Shirdi.

D:Drive 2021-G-Quotation-Letter-Head-P.No 12



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014
Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.
Website: <http://www.shrisaibabasansthan.org> E_mail: saibaba_anr@sancharnetin/saisandesh_anr@sancharnet.in