



# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post Shirdi - 431109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India)  
Telephone: 91 - 2423 - 238 671 672 673, Fax: 91 - 2423 - 238 660, Telegram: Saibab, Shirdi  
Email: saibabashospital@gmail.com

NO. 559/W5/SSM/STORES/(P/S)/ 462 /2022

Date: 13 JUN 2022

To,

**Sub: - Quotation for Purchase of Dental Chair Unit with Light Cure Unit and Ultrasonic Scalar against buyback of old Dental Chair for Dental Department of Shri Saibaba Hospital, Shirdi.**

Dear Sir,

We have to purchase Dental Chair Unit with Light Cure Unit and Ultrasonic Scalar against buyback of old Dental Chair for Dental Department of Shri Saibaba Hospital, Shirdi given in enclosed Annexure "A" & the following terms and conditions.

## Terms and Condition

01. The Dental Chair Unit with Light Cure Unit and Ultrasonic Scalar will have to be supplied F.O.R. Shirdi.
02. The rate should be inclusive of GST, transport charges, T.O.T. Etc.
03. You will have to mention name of the company, packing, rate per unit, total cost etc.
04. Packing, forwarding, and other charges will not be paid by the Sansthan.
05. We have to buyback old Dental Chair (Make- Confident, Model - Chamundi) so please mention the cost of it.
06. The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
07. In case if supply is not made in the proper time Sansthan will free to purchase the concern material from other party of open market and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
08. You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
09. In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.

Page No. 2/-

2021-G Quotation Letter Head P No 7



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014  
Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798, Telegram: Sainiketan, Mumbai.  
Website: <http://www.shrisaibasansthan.org>, E-mail: saibaba\_mum@saibabamumbai.org/saibabamumbai@gmail.com

श्री साईबाबा संस्थान रुग्णालये, शिर्डी





# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post Shirdi - 423109, Tal. Patana, Dist. Ahmednagar, Maharashtra (India)  
Telephone: +91 - 2423 - 258 671 472 473, Fax: +91 - 2423 - 258 660, Telegram: Sainiketan, Mumbai.  
Email: saibabashospital@yahoo.co.in

Page No. 2

10. Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
11. All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the award nor after the award for any part or accessory.
12. Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.
13. The supplier has to mention the one time visit charges for breakdown and servicing.
14. Equipment mentioned in the Technical specification shall operate on 230V  $\pm 10\%$ , 50Hz, single phase electric supply if not mentioned specifically.
15. The necessary protective relaying / circuitry shall be there with the machines.
16. Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to SSI officials as mentioned elsewhere in the tender specifications.
17. The product should be time tested, branded and should be OEM.
18. The life of the equipment should be mentioned.
19. The Supplier must quote details of Technical Specification along with coloured brochure. In absence of these details, quotation will not be considered.
20. The quotation must be clearly mention on the sealed envelope "Quotation for Dental Chair Unit with Light Cure Unit and Ultrasonic Scalar against buyback of old Dental Chair for Dental Department of Shri Saibaba Hospital".
21. The quotation in sealed cover may please be submitted to our Shirdi office on or before Dt. 20/08/2022.

Thanking you and with blessings of Shri Saibaba.

Yours faithfully,

Medical Director,  
Shri Saibaba Hospital, Shirdi





# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba.hospital@sai.org.in

## Shri Saibaba Hospital, Shirdi.


### Annexure "A"

#### Requirement of Dental Chair against buyback of old Dental Chair for Dental Department of Shri Saibaba Hospital.

(Preferred Make - Confident, Suzdent, Gnatus, Roodenta, Promodent or equivalent)

Sr. No.	Name of Item	Qty.	Make	Rate Per in Rs.	Total Amt. in Rs.
1	Dental Chair Unit with Light Cure Unit and Ultrasonic Scalar	2			
	· Should be electrically operated with Multi programmable for both left and right handed dentistry option				
	· Dental chair should have erasable programs with microprocessor controlled where user can set his own programs. The program switch should be fixed to the instrument tray. Program 0 and Gargling 1 & 2 erasable program.				
	· Chair should perfectly body contoured. Assistant console for chair control.				
	· The right arm side of the chair should have lateral rotation for easy access of the patient.				
	· Should have a On/Off sensor switch non touch operated operating light (Halogen/LED light) with imported glass reflector with minimum two steps of intensity control from 25000 -40000 Lux with 3 axis movements.				
	· Should have auto water connection for spittoon and tumbler with auto fillingsensor.				
	· Should have Ceramic spittoon & tumbler-cup. It should be single piece high quality material with perfect smooth and with removable and autoclavable cannula for better hygiene.				
	· Should have high and low motorized vacuum suction with flow control valve				
	· Should have Saliva ejector with low				

2021-G-Quotation-Letter-Head-P.No.3



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014  
Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798, Telegram: Sainiketan, Mumbai.  
Website: <http://www.shrisaibabasansthan.org>, E-mail: [saibaba\\_anr@sancharnet.in](mailto:saibaba_anr@sancharnet.in), [saibaba\\_anr@saisandesh.in](mailto:saibaba_anr@saisandesh.in), [saibaba\\_anr@sancharnet.in](mailto:saibaba_anr@sancharnet.in)





PTR. NO. : E-69, AHMED NAGAR

**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba.hospital@sai.org.in

Sr. No.	Name of Item	Qty.	Make	Rate Per in Rs.	Total Amt. in Rs.
	suction				
	Should have an auto drain and auto flush system,				
	Water-cup filler should be operated by switch				
	Should have a multi functional foot control				
	Should have an SS Large instrument tray & adjustable headrest and an LED x-ray viewer				
	Modular delivery system over the Patient. Should have two air rotor points with hand piece and micro motor 35,000 RPM with endo check & niti files display of speed with handpiece and two sets 3 way syringe.				
	Should provide Dental operator's stool with back rest movement and hand rest.				
	Should be supplied with medical grade, noiseless compressor of 0.75 HP or more.				
	Should be operated in input supply 200 to 240 V, AC, 50Hz.				
	Should be supplied with Dental materials starter kit (CONSUMABLE)				
	Accessories:				
	1. Air rotor Unit with control box fitted with transparent regulated water tank, multifunctional foot control. Supplied with Titanium/SS Super torque airtor hand piece with cellular optic light , Ultra push type non retractive valve(NSK/Sirona/W&H)	2			
	2. Air rotor Handpiece clean head with speed o 3,50,000 RPM (NSK/Sirona/W&H)	2			
	3. Contra angle Handpiece Autoclavable with speed upto 40,000 RPM (NSK/Sirona/W&H)	2			

2021-G-Quotation-Letter-Head-P.No 4



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014  
Telephone: +91 - 22 - 24166556, Fax: +91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.  
Website: <http://www.shrisaibabasansthan.org>, E\_mail: [saibaba\\_anr@sancharnet.in](mailto:saibaba_anr@sancharnet.in), [saisandesh\\_anr@sancharnet.in](mailto:saisandesh_anr@sancharnet.in)

# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

PTR. NO. : E-69, AHMED NAGAR

SHIRDI Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba.hospital@aunl.org.in



Sr. No.	Name of Item	Qty.	Make	Rate Per In Rs.	Total Amt. in Rs.
4.	Straight hand piece Autoclavable with speed upto 40,000 RPM (NSK/Sirona/W&H)	2			
5.	Brushless Micromotor with speed 2000-40,000 RPM wit digital display of speed (NSK/Sirona/W&H)	2			
6.	Oil free, Medical grade, minimum/no noise compressor of 0.75 HP-1 HP. It should be fitted with thermo cut-off relay system, Auto head air release valve, drain valve, pressure gauge. The inner surface should be coating with epoxy to prevent rusting	2			
7.	Multi function foot control,	2			
8.	LED X- ray viewer,	2			
9.	Stainless steel instrument tray,	2			
10.	Dental Operators stool (with foot ring)	2			
11.	Forceps kit = Set of 12.	2			
12.	Root elevator= Set of 6	2			
13.	Small instrument (pmt & filling ins. Etc.)	2			
14.	3-way syringe for Air, Water & spray,	2			
15.	Supreme optic scalar unit with scaling tips: Piezotronic scalar 28000/ 32000 KHz frequency. Autoclavable Handpiece total control is micro processor based control unit. Hand pieces with sleek design. The scalar should be supplied with Piezotronic Scalar with 4 Nos. of tips (EMS/Satelac/NSK/W&H)	2			
16.	Beecool turbo ILED Light cure unit: ensures upto 1300mW/sq.cm with 4 curing modes (RAMP, Fast, Long, Pulse). It should be supplied with Li-ion battery with sufficient capacity. a) It can work upto 4000 sec.; b) Wavelength of 470 nm ;c) Full color LCD display	2			
17.	Dental materials starter kit (CONSUMABLE)	2			
	Drainage system should be done by the company as turnkey work.				

2021-G-Quotation-Letter-I Head-P.No.5



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014  
Telephone: +91 - 22 - 24166556, Fax: +91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.  
Website: <http://www.shrisaibabasansthan.org>, E-mail: [saibaba\\_aunl@saichanctin/saisandesh\\_aunl@saichanctin](mailto:saibaba_aunl@saichanctin/saisandesh_aunl@saichanctin)






# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 433109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India)  
Telephone: +91 - 2423 - 238 671-672-673, Fax: +91 - 2423 - 238660, Telegram: Sainath, Shirdi  
Email: saibaba@hospitalshirdi.org

Sr. No.	Name of Item	Qty.	Make	Rate Per In Rs.	Total Amt. In Rs.
	· All Technology should be mandatorily with safety approval - FDA / UL and CE. Approval certificate is to be submitted in addition to CE Certificate for the total unit.				
	· The product should be time tested, branded and should be OEM				
	· Warranty- 1 Years				
	· LAMC Cost- AMC cost should be including all Taxes and it should be fixed for next 5 years.				
	· Warranty and LAMC should including 4 PMS visits and unlimited Breakdown calls & Calibration whenever necessary free of cost.				
	· List & Cost of Consumables to be fixed for warranty period.				
	Buyback Offer for old machine Make- Confident, Model - Chamundi	2 Nos.			
	Net Amt. in Rs.				

  
Medical Director,  
Shri Saibaba Sansthan Trust,  
Hospitals, Shirdi

