PTR. NO.: E-69, AHMED NAGAR



### SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).

Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 · 2423 - 258660, Telegram: Sainath, Shirdi.

Email: saibaba•hospital@sai•org•in

	NO.SSS/WS/SSBH/STORES/(P/S)/ DATE:	/2018
Го,		
	·	
	·	

### <u>Sub: - Quotation for Purchase of Vertical Autoclave Machine</u> for Pathology Department of Shri Saibaba Hospital, Shirdi.

#### Dear Sir,

We have to purchase Vertical Autoclave Machine for Shri Saibaba Hospital, Shirdi given in enclosed Annexure "A" & the following terms and conditions.

### **Terms and Condition**

- 1. The **Vertical Autoclave Machine** will have to be supplied F.O.R. Shirdi.
- 2. The rate should be inclusive of GST, transport charges, T.O.T. Etc.
- 3. You will have to mention name of the company, packing, rate per unit, total cost etc.
- 4. Packing, forwarding, and other charges will not be paid by the Sansthan.
- 5. The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
- 6. In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
- 7. You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
- 8. In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
- 9. Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
- 10. All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the wards nor after the award for any part or accessory.
- 11. Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.

Page No. 2/-



PTR. NO. : E-69. AHMED NAGAR



## SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).

Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 · 2423 - 258660, Telegram: Sainath, Shirdi.

Email: saibaba•hospital@sai•org•in

#### Page No. 2


- 12. The supplier has to mention the one time visit charges for breakdown and servicing.
- 13. Equipment mentioned in the Technical specification shall operate on 230V ±10%, 50Hz, single phase electric supply if not mentioned specifically.
- 14. The necessary protective relaying / circuitry shall be there with the machines.
- 15. The equipment shall have valid CE marking and US FDA approval (Wherever necessary) and documentary evidence to that effect shall be submitted.
- 16. The equipment shall be having warranty of 01 year. The warranty and CMC shall cover the batteries and lamps irrespective of whether those are treated as consumables or otherwise.
- 17. Minimum 4 Nos. of preventive maintenance visits and unlimited breakdown call visits as per requirement must be attended in the warranty and AMC/CMC period. The services to the equipment in warranty and CMC period will be considered as same.
- 18. The equipment should be provided with one hard copy in original of the detailed service manual and operation manual. Further, a soft copy is also required.
- 19. Repairs and maintenance training to two Biomedical Engineers should be given for one week or more as per requirement.
- 20. The equipment must be tropicalized as below-

Working Temperature : Max.45<sup>o</sup> Storage Temperature : Max.50<sup>o</sup>C,

Relative humidity for working : 40 – 60%, Relative humidity for Storage : Max 90%

- 21. Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to SSST officials as mentioned elsewhere in the tender specifications.
- 22. The supplier has to submit user list with contact telephone numbers for the model quoted.
- 23. The supplier has to give past performance certificate of the quoted equipment.
- 24. The supplier should submit all the Datasheets of the quoted model for technical comparison.
- 25. The product should be time tested, branded and should be OEM.

Page No. 3/-



PTR. NO. : E-69. AHMED NAGAR



### SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 · 2423 - 258660, Telegram: Sainath, Shirdi.
Email: saibaba•hospital@sai•org•in

#### Page No. 3

- 26. The life of the equipment should be mentioned.
- 27. The quotation must be clearly mention on the sealed envelope <u>"Vertical Autoclave Machine of Pathology Department for Shri Saibaba Hospital"</u>.
- 28. The quotation in sealed cover may please be submitted to our Shirdi office on or before **Dt. 15/08/2018.**

Thanking you and with blessings of Shri Saibaba.

Yours Sincerely,

Administrative Officer Shri Saibaba Sansthan Trust, Shirdi



PTR. NO.: E-69, AHMED NAGAR



## SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 · 2423 - 258660, Telegram: Sainath, Shirdi.
Email: saibaba•hospital@sai•org•in

### Shri Saibaba Hospital, Shirdi Annexure "A"

Technical Specification for Vertical Autoclave Machine - Pathology Department (Preferred Make: - M. shah & Co., Modi, Pathak, Labtop or equivalent)

Sr.	Description	Qty.	Rate Per	Total
No.	_		Inclu. of	Amt. in
			All Tax	Rs.
1	VERTICAL AUTOCLAVE	01		
	It should meet the complete sterilization			
	requirements including of instruments, glassware,			
	liquids, culture media, pipette tips, contaminated			
	media, biological waste & suitable for use in			
	laboratories.			
	Construction			
	All Stainless steel Double wall construction,			
	Inside SS 304 & outside SS 316			
	Top Lid should be thick SS 304 die pressed in			
	mirror finish and fitted to the Body by means of			
	Silicone rubber gasket.			
	The unit should be fitted with heavy locking			
	system (flange fly 7/8 wing nuts/single lever) to			
	hold the lid at high pressure.			
	• Unit should be fitted with Main Pressure			
	Gauge/indicator, Pressure Control Valve, Steam			
	Release Cock/ exhaust valve, drain valve and			
	extra safety valve to provide safety to the users.			
	Heating should be done by Kettle Type			
	immersion heating elements made of high grade			
	material and duly ISI Marked			
	Automatic Sterilizing cycle at 121°-122°C.			
	Brass case Thermometer with arrangement			
	Material Lifting Arrangement (Foot Pedal/any			
	other suitable) should be available			
	Water level Indicating Gauge- Glass/any other			
	type			
	Automatic pressure control switch should be			
	provided			
	Automatic low water level cut off device should			
	be provided			



PTR. NO.: E-69, AHMED NAGAR



# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).

Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 · 2423 - 258660, Telegram: Sainath, Shirdi.

Email: saibaba•hospital@sai•org•in

Sr. No.	Description	Qty.	Rate Per Inclu. of All Tax	Total Amt. in Rs.
	<ul> <li>It should have fast cooling system</li> <li>A perforated Stainless Steel Basket should be provided</li> <li>Depressurizing &amp; Unlocking the unit should be manual</li> <li>It should be provided with accessory support comprising digital timer, vacuum breaker, Drain Valve</li> <li>Works on 200-230 V AC, 50 Hz Single phase supply</li> <li>Working Pressure &amp; Temperature         <ul> <li>The working pressure for sterilization process should be 15 psi -20 psi.</li> <li>The working Temperature for sterilization process should be 121°-122°C.</li> </ul> </li> <li>Capacity: min. 75 Ltrs         <ul> <li>All Technology should be mandatorily with safety approval – CE/ BIS Approved. Approval certificate is to be submitted in addition to CE Certificate for the total unit.</li> <li>The product should be time tested, branded and should be OEM.</li> <li>Warranty period shall be 1 year.</li> <li>AMC / CMC cost: AMC / CMC cost should be including Taxes. It should be fixed for next 5 years after warranty.</li> <li>Warranty, AMC, CMC should include PMS VISITS-4 Nos. &amp; unlimited Breakdown calls with Calibration whenever necessary free of cost. In addition, Warranty &amp; CMC includes all spare parts.</li> <li>Cost of all types of consumables &amp; Accessories should be given for next five years inclusive all</li> </ul> </li> </ul>		All Tax	Rs.
	Taxes.			

**Medical Director,** Shri Saibaba Hospitals, Shirdi

