

**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).

Telephone: +91 - 2423 - 258 670-671-672. Fax: +91 - 2423 - 258676.

Email : hospital.purchase@sai.org.in/ hospital.store@sai.org.in

NO.SSS/WS/SSBH/STORES/(P/S)/803 /2022

DATE:

16 JUL 2022

To,

**SUB: - Quotation for Urine Analyser Machine for Laboratory Dept. of
Shri Sainath Hospital.**

Dear Sir,

We have to purchase of Urine Analyser Machine for Shri Sainath Hospital, Shirdi as per following Technical Specification & the following terms and conditions.

Sr.No.	Description	Qty.	Rate Per in Rs.	Total Amt.in Rs.
1.	<p><u>TECHNICAL SPECIFICATIONS for Urine Analyser</u></p> <p>1. Principal -semi-automated semi-quantitative reflex photometer.</p> <p>2. Instrument Have measuring using a CCD camera/LED.</p> <p>3. Measuring Range between wave lengths: 535, 610 nm.</p> <p>4. Latency: 60 s</p> <p>5. Capacity: 400 strips per hour</p> <p>6. Touch screen (LCD display): 5"(320x240) colour TFT</p> <p>7. user-defined setting option.</p> <p>8. Built-in thermo printer, print speed 12 results per minute</p> <p>9. External communication via RS 232 interface</p> <p>10. Optional connection to a PC, an external keyboard or a barcode scanner</p> <p>11. MemoryCapacity: Last 1000 measurements.</p> <p>12. Operators should be given training for its regular use & maintenance training should be given to Biomedical Engineer free of cost.</p> <p>13. All technology should be mandatorily with safety approval-US FDA/UL/BIS & CE Approval</p>	1		

M:/Quotation Letter E-Drive



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014

Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.

Website: <http://www.shrisaibabasansthan.org>, E-mail: saibaba_mum@sancharnet.in/saisandesh_mum@sancharnet.in

**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).


Telephone: +91 - 2423 - 258 670-671-672. Fax: +91 - 2423 - 258676.

Email : hospital.purchase@sai-org-in/ hospital.store@sai-org-in

Sr.No.	Description	Qty.	Rate Per in Rs.	Total Amt.in Rs.
	certificate must be submitted for the total unit. 14. The product should be time tested, branded & should be OEM. 15. Warranty -05 year.It should include PMS Visits-4 Nos.& unlimited breakdown calls & calibration with all spare parts whenever necessary. 16. CAMC-CAMC cost should be quoted inclusive all Taxes for next 5 years. 17. Cost of consumables to be fixed for warranty period.			

Terms and Conditions

01. The Urine Analyser Machine will have to be supplied F.O.R. Shirdi.
02. The rate should be inclusive of GST, transport charges, T.O.T. Etc.
03. You will have to mention name of the company, packing, rate per unit, total cost etc.
04. Packing, forwarding, and other charges will not be paid by the Sansthan.
05. The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
06. In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
07. You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
08. In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
09. Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
10. All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the wards nor after the award for any part or accessory.
11. Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.
12. The supplier has to mention the one time visit charges for breakdown and servicing.



- Thanking you and with blessings of Shri Saibaba,

Yours faithfully,

Medical Director,
Shri Sai Baba Hospital, Shirur

