

**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba•hospital@sai•org•in

NO.SSS/WS/SSBH/STORES/(P/S)/

/2017

DATE:

To,

.....  
.....  
.....

**Sub : - Quotation for Purchase of Arthroscope 30 Deg., 4 mm for  
Shri Saibaba Hospital, Shirdi.**

Dear Sir,

We have to purchase Arthroscope 30 Deg., 4 mm for Shri Saibaba Hospital, Shirdi given in enclosed Annexure "A" & the following terms and conditions.

**Terms and Condition**

1. The Arthroscope 30 Deg., 4 mm will have to be supplied F.O.R. Shirdi.
2. The rate should be inclusive of GST, transport charges, T.O.T. Etc.
3. You will have to mention name of the company, packing, rate per unit, total cost etc.
4. Packing, forwarding, and other charges will not be paid by the Sansthan.
5. The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
6. In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
7. You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
8. In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
9. Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
10. All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the wards nor after the award for any part or accessory.
11. Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.
12. The supplier has to mention the one time visit charges for breakdown and servicing.

Page No. 2/-





# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba•hospital@sai•org•in

Page No. 2

13. Equipment mentioned in the Technical specification shall operate on 230V  $\pm$ 10%, 50Hz, single phase electric supply if not mentioned specifically.
14. The necessary protective relaying / circuitry shall be there with the machines.
15. The equipment shall have valid CE marking and US FDA approval (Wherever necessary) and documentary evidence to that effect shall be submitted.
16. The equipment shall be having warranty of 01 year. The warranty and CMC shall cover the batteries and lamps irrespective of whether those are treated as consumables or otherwise.
17. Minimum 4 Nos. of preventive maintenance visits and unlimited breakdown call visits as per requirement must be attended in the warranty and AMC/CMC period. The services to the equipment in warranty and CMC period will be considered as same.
18. The equipment should be provided with one hard copy in original of the detailed service manual and operation manual. Further, a soft copy is also required.
19. Repairs and maintenance training to two Biomedical Engineers should be given for one week or more as per requirement.
20. The equipment must be tropicalized as below- Working Temperature : Max.45<sup>o</sup>  
Storage Temperature : Max.50<sup>o</sup>C, Relative humidity for working : 40 - 60%,  
Relative humidity for Storage : Max 90%
21. Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to SSST officials as mentioned elsewhere in the tender specifications.
22. The supplier has to submit user list with contact telephone numbers for the model quoted.
23. The supplier has to give past performance certificate of the quoted equipment.
24. The supplier should submit all the Datasheets of the quoted model for technical comparison.
25. The product should be time tested, branded and should be OEM.
26. The life of the equipment should be mentioned.
27. The quotation must be clearly mention on the sealed envelops "Arthroscope 30Deg., 4mm of Shri Saibaba Hospital".
- 29) The quotation in sealed cover may please be submitted to our Shirdi office on or before Dt. 26/10/2017.

Thanking you and with blessings of Shri Saibaba.

Yours Sincerely,

Administrative Officer,  
Shri Saibaba Sansthan Trust, Shirdi.



**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba•hospital@sai•org•in

**Shri Saibaba Hospital, Shirdi****Annexure "A"****Technical Specification for Arthroscope**

Sr. No.	Description	Qty.	Rate Per Incl. of All Tax	Total Amt. in Rs.
1	<b>Arthroscope 30 Deg., 4 mm</b>	1		
	· It should have Diameter 4mm, tip angle 30 degree & working length min. 150 mm			
	· It should be Autoclavable			
	· Tip lens should be scratch free			
	· Optics should be High quality multicoated (anti reflective) and of latest generation, Highly refractive glass for high resolution brightness of image			
	· Optical system should have better clarity and depth of focus			
	· Optical system and mechanical system should have extremely precise centering so that the image should be clear, bright with high contrast upto the edges			
	· It should be compatible with the Karl Storz Arthroscopy shaver set			
	· Tube design should be such that it should have very high stability with all types of endoscopy instruments			
	· All joints should be laser welded so that there should not be any leakage			
	· Care should be taken to avoid interior condensation.			
	· All Technology should be mandatorily with safety approval - FDA, CE, UL or BIS approved. Approval certificate is to be submitted for the total unit.			




**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba-hospital@sai-org.in

Page No.2

Sr. No.	Description	Qty.	Rate Per Inclu. of All Tax	Total Amt. in Rs.
	· Manufacturer/Supplier should have ISO certification for quality standards.			
	· The product should be time tested, branded and should be OEM			
	· <b>Warranty- 1 Year</b>			
	<b>After the warranty period is over, Annual Comprehensive maintenance Contract (CMC) cost –</b> CMC cost should be quoted as per your basis including Taxes. It should be fixed for next 5 years i.e. from 2 <sup>nd</sup> to 6 <sup>th</sup> year after completion of warranty period. The Successful supplier has to ensure that all the required spares and services are available free of cost during the period of Warranty and CMC period.			
	<b>Cost of Consumables to be fixed for warranty period.</b>			

  
Medical Director,  
Shri Saibaba Sansthan Trust,  
Hospitals, Shirdi

