



# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba•hospital@sai•org•in

NO.SSS/WS/SSBH/STORES/(P/S)/ /2018  
DATE:

To,

-----  
-----  
-----  
-----

## Sub : - Quotation for Purchase of Catract Set for Shri Saibaba Hospital, Shirdi.

Dear Sir,

We have to purchase of Cactract Set for Shri Saibaba Hospital, Shirdi given in enclosed Annexure "A" & the following terms and conditions.

### Terms & Condition

- 01) 01) The Cactract Set will have to be supplied F.O.R. Shirdi.
- 02) The rate should be inclusive of GST, transport charges, T.O.T. Etc.
- 03) You will have to mention name of the company, packing, rate per unit, total cost etc.
- 04) Packing, forwarding, and other charges will not be paid by the Sansthan.
- 05) The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
- 06) Please note that if taxes are not mentioned then rates quoted will be considered as including all taxes & F.O.R. Shirdi & If we demand the rates with AMC /CMC & supplier fails to give rates with AMC / CMC then concern quotation may liable for rejection.
- 07) In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
- 08) You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
- 09) In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
- 10) Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.

Page No. 2/-





# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

**SHIRDI:** Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
**Telephone:** +91 - 2423 - 258 671-672-673, **Fax:** +91 -2423 - 258660, **Telegram:** Sainath, Shirdi.  
**Email :** saibaba•hospital@sai•org•in

Page No.2

- 11) All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the wards nor after the award for any part or accessory.
- 12) Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.
- 13) The supplier has to mention the one time visit charges for breakdown and servicing.
- 14) Equipment mentioned in the Technical specification shall operate on 230V  $\pm$ 10%, 50Hz, single phase electric supply if not mentioned specifically.
- 15) The necessary protective relaying / circuitry shall be there with the machines.
- 16) The equipment shall have valid CE marking and US FDA approval (Wherever necessary) and documentary evidence to that effect shall be submitted.
- 17) The equipment shall be having warranty of 01 year. The warranty and CMC shall cover the batteries and lamps irrespective of whether those are treated as consumables or otherwise.
- 18) Minimum 4 Nos. of preventive maintenance visits and unlimited breakdown call visits as per requirement must be attended in the warranty and AMC/CMC period. The services to the equipment in warranty and CMC period will be considered as same.
- 19) The equipment should be provided with one hard copy in original of the detailed service manual and operation manual. Further, a soft copy is also required.
- 20) Repairs and maintenance training to two Biomedical Engineers should be given for one week or more as per requirement.
- 21) The equipment must be tropicalized as below-  
Working Temperature : Max.450C  
Storage Temperature : Max.500C  
Relative humidity for working : 40 - 60%  
Relative humidity for Storage : Max 90%
- 22) Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to SSST officials as mentioned elsewhere in the tender specifications.
- 23) The supplier has to submit user list with contact telephone numbers for the model quoted.
- 24) The supplier has to give past performance certificate of the quoted equipment.

Page No. 3/-





## **SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

**SHIRDI:** Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
**Telephone:** +91 - 2423 - 258 671-672-673, **Fax:** +91 -2423 - 258660, **Telegram:** Sainath, Shirdi.  
**Email :** saibaba•hospital@sai•org•in

Page No.3

- 25) The supplier should submit all the Datasheets of the quoted model for technical comparison.
- 26) The supplier should submit all the Datasheets with Coloured Browcher of the quoted model for technical comparison.
- 27) The product should be time tested, branded and should be OEM.
- 28) The life of the equipment should be mentioned.
- 29) The quotation must be clearly mention on the sealed envelops “ Cattract Set “ for Shri Saibaba Hospital.
- 30) The quotation in sealed cover may please be submitted to our Shirdi office on or before Dt.25/08/2018.

Thanking you and with blessings of Shri Saibaba

Yours Sincerely,

Administrative Officer,  
Shri Saibaba Sansthan Trust's  
Shirdi





PTR. NO. : E-69, AHMED NAGAR

# SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 -2423 - 258660, Telegram: Sainath, Shirdi.  
Email : saibaba•hospital@sai•org•in

## Shri Saibaba Hospital, Shirdi. Annexure "A"

Sr.No.	Name of Instrument	Qty.	Rate per.	Total Amt.
<b>A</b>	<b>Catract set</b>			
1	Barraquer wire speculum	10		
2	Phaco speculum	10		
3	Dastoor iris repositor	10		
4	Wire Vectis	10		
5	Sinsky Hook R/D Angled	10		
6	Castroviejo Caliper 20mm	10		
7	Masquito Forceps Straight	20		
8	Mosquito Forceps Curved	20		
9	Towel Clip	10		
10	Dastoor Superior Rectus	10		
12	Jaffe tying Forceps Straight	10		
13	Plaufique Suture Forceps	10		
14	Lims Suture Forceps	10		
15	Kelman Mcpherson 11mm	10		
16	Iris Scissor, Curved	10		
17	Vannas Scissor Straight	10		
18	Conjunctival Scissors	10		
19	Corneal Scissors Small Blade	10		
20	Capsulotomy Scissors 13mm	10		
21	Fine Needle Holder Curved	10		

2010-11-T/Quotation-2010/4



MUMBAI: Sainiketan, 804 - B, Dr. Ambedkar Marg, Dadar, Mumbai - 400014  
Telephone: + 91 - 22 - 24166556, Fax: + 91 - 22 - 24150798. Telegram: Sainiketan, Mumbai.  
Website: <http://www.shrisaibabasansthan.org>, E\_mail: [saibaba\\_anr@sancharnetin/saisandesh](mailto:suibaba_anr@sancharnetin/saisandesh), [anr@sancharnet.in](mailto:anr@sancharnet.in)



PTR. NO. : E-69, AHMED NAGAR

## **SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

**SHIRDI:** Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).  
**Telephone:** +91 - 2423 - 258 671-672-673, **Fax:** +91 -2423 - 258660, **Telegram:** Sainath, Shirdi.  
**Email :** saibaba•hospital@sai•org•in

22	Irrigating Aspirating Cannula	10		
23	Air Cannula 27G	20		
24	Air Cannula	20		
25	Instrument Tray with 1 Silicon Mat	10		
26	B.P. Handle No.3	10		
27	Barraquer Needle Holder W/O Lock	10		
28	Capsulorhexis Forceps	10		
29	8/0 Suture Tying Forceps	10		
30	Aluminium Box With mat	10		
31	AC Maintainer	10		
32	Aspiration Cannula 22G	10		
<b>B</b>	<b>Self Retaining Eye Speculum</b>	05		

**Medical Director**  
**Shri Saibaba Sansthan Trust's**  
**Hospital, Shirdi**

