



SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba•hospital@sai•org•in

NO.SSS/WS/SSBH/STORES/(P/S)/ /2018

DATE:

To,

SUB: - Quotation for Purchase of Hot Air Oven of Shri Saibaba Hospital.

Dear Sir,

We have to Hot Air Oven For Pathology Dept. of Shri Saibaba Hospital, Shirdi given in enclosed Annexure "A" & the following terms and conditions.

Terms & Condition

- 01) The Hot Air Oven will have to be supplied F.O.R. Shirdi.
- 02) The rate should be inclusive of GST, transport charges, T.O.T. Etc.
- 03) You will have to mention name of the company, packing, rate per unit, total cost etc.
- 04) Packing, forwarding, and other charges will not be paid by the Sansthan.
- 05) The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
- 06) Please note that if taxes are not mentioned then rates quoted will be considered as including all taxes & F.O.R. Shirdi.
- 07) In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
- 08) You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
- 09) In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
- 10) Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
- 11) All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the wards nor after the award for any part or accessory.
- 12) Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.

Page No. 2/-





SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, **Fax:** +91 -2423 - 258660, **Telegram:** Sainath, Shirdi.
Email : saibaba•hospital@sai•org•in

Page No.2

- 13) The supplier has to mention the one time visit charges for breakdown and servicing.
- 14) Equipment mentioned in the Technical specification shall operate on 230V $\pm 10\%$, 50Hz, single phase electric supply if not mentioned specifically.
- 15) The necessary protective relaying / circuitry shall be there with the machines.
- 16) The equipment shall be having warranty of 01 year. The warranty and CMC shall cover the batteries and lamps irrespective of whether those are treated as consumables or otherwise.
- 17) Minimum 4 Nos. of preventive maintenance visits and unlimited breakdown call visits as per requirement must be attended in the warranty and AMC/CMC period. The services to the equipment in warranty and CMC period will be considered as same.
- 18) The equipment should be provided with one hard copy in original of the detailed service manual and operation manual. Further, a soft copy is also required.
- 19) Repairs and maintenance training to two Biomedical Engineers should be given for one week or more as per requirement.
- 20) The equipment must be tropicalized as below-
Working Temperature : Max.450C
Storage Temperature : Max.500C
Relative humidity for working : 40 - 60%
Relative humidity for Storage : Max 90%
- 21) Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to SSST officials as mentioned elsewhere in the tender specifications.
- 22) The product should be time tested, branded and should be OEM.
- 23) The life of the equipment should be mentioned.
- 24) The quotation must be clearly mention on the sealed envelops "Hot Air Oven" for Shri Saibaba Hospital.
- 25) The quotation in sealed cover may please be submitted to our Shirdi office on or before Dt. 26/10/2018.

Thanking you and with blessings of Shri Saibaba.

Yours Sincerely,

Administrative Officer
Shri Saibaba Sansthan Trust
Shirdi



**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 -2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba•hospital@sai•org•in

Shri Saibaba Hospital, Shirdi.**Annexure "A"****Technical Specification of Hot Air Oven for Shri Saibaba Hospital.**

Sr. No.	Description	Qty.	Rate Per in Rs. (Incl. of all Taxes & Duties)	Total Amt. in Rs.
01	<ul style="list-style-type: none">• Electronic Controlled Sterilizing Process	02		
	<ul style="list-style-type: none">• Outer case and internal chamber – constructed from stainless steel.			
	<ul style="list-style-type: none">• Two or three adjustable shelves.			
	<ul style="list-style-type: none">• Wall and hinged Door Insulated with High Grade Glass Wool.			
	<ul style="list-style-type: none">• Door with Lock and Flat Key.			
	<ul style="list-style-type: none">• Temperature: adjustable from +50°C to +200 °C ±10 %.			
	<ul style="list-style-type: none">• Timer Range: 0-250 minutes. ±10 %.			
	<ul style="list-style-type: none">• Capacity: 50- 60l.			
	<ul style="list-style-type: none">• Power requirements			
	<ul style="list-style-type: none">• 20 VAC ±10 %, 50Hz.			
	<ul style="list-style-type: none">• Warranty-One year Warranty.			
	<ul style="list-style-type: none">• Il technology should be mandatorily with safety approval. Approval certificate is to be submitted in addition to CE Certificate for the total unit.			
	<ul style="list-style-type: none">• he product should be time tested, branded & should be OEM.			
	<ul style="list-style-type: none">• AMC/CMC Cost should be including Taxes. It should be fixed for next 5 Years.			
	<ul style="list-style-type: none">• Cost of Consumables to be fixed for warranty period.			
			Total Amt.in Rs.	

**Medical Director
Shri Saibaba Sansthan Trust's
Hospital, Shirdi**

