

**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India)
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi
Email : saibaba•hospital@sai•org•in

NO.SSS/WS/SSBH/STORES/(P/S)/ 195 /2022

DATE: 28 APR 2022

To,

Sub : - Quotation for Purchase of 10 KVA UPS for Ortho O.T. of Shri Saibaba Hospital, Shirdi.

Dear Sir,

We have to purchase 10 KVA UPS for our Shri Saibaba Hospital, Shirdi given in enclosed Annexure "A" & the following terms and conditions.

Terms & Condition

- 01) The 10 KVA UPS will have to be supplied F.O.R. Shirdi.
02. The rate should be inclusive of GST, transport charges, T.O.T. Etc.
03. You will have to mention name of the company, packing, rate per unit, total cost etc.
04. Packing, forwarding, and other charges will not be paid by the Sansthan.
05. The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
06. In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
07. You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
08. In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
09. Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
10. All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the wards nor after the award for any part or accessory.

Page No.2/-



**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi
Email : saibaba-hospital@sai-org-in

Page No. 2

11. Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.
12. The supplier has to mention the one time visit charges for breakdown and servicing.
13. Equipment mentioned in the Technical specification shall operate on 230V \pm 10%, 50Hz, single phase electric supply if not mentioned specifically.
The necessary protective relaying / circuitry shall be there with the machines.
14. The equipment shall have valid CE marking IEC approval (Wherever necessary) and documentary evidence to that effect shall be submitted.
15. The equipment shall be having warranty of one year. The warranty and AMC/CMC shall cover the batteries and lamps irrespective of whether those are treated as consumables or otherwise.
16. Minimum 4 Nos. of preventive maintenance visits and unlimited breakdown call visits as per requirement must be attended in the warranty and AMC/CMC period.
17. The services to the equipment in warranty and AMC/CMC period will be considered as same.
18. The equipment should be provided with one hard copy in original of the detailed service manual and operation manual. Further, a soft copy is also required.
19. Repairs and maintenance training to two Biomedical Engineers should be given for one week or more as per requirement.
20. The equipment must be tropicalized as below-
Working Temperature : Max.45^o
Storage Temperature : Max.50^oC,
Relative humidity for working : 40 - 60%,
Relative humidity for Storage : Max 90%
21. Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to SSST officials as mentioned elsewhere in the tender specifications.
22. The supplier has to submit user list with contact telephone numbers for the model quoted.
23. The supplier has to give past performance certificate of the quoted equipment.

Page No. 3/-



**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba-hospital@sai-org-in

Page No. 3

24. The supplier should submit all the Datasheets of the quoted model for technical comparison.
25. The product should be time tested, branded and should be OEM.
26. The life of the equipment should be mentioned.
27. The Supplier must quote details of Technical Specification along with coloured brochure. In absence of these details, quotation will not be considered.
28. The quotation must be clearly mention on the sealed envelope "**Quotation for "10 KVA UPS "for Shri Saibaba Hospital"**".
29. The quotation in sealed cover may please be submitted to our Shirdi office on or before Dt. 14/05/2022.
- 30) In case of any dispute the same will be subject to the jurisdiction of the competent court at Rahata / Kopergaon.

Thanking you and with blessings of Shri Saibaba.

Yours sincerely,

Medical Director,
Shri Saibaba Sansthan Trust
Hospital, Shirdi



**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post. Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India)
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba-hospital@sai-org-in

Shri Saibaba Hospital, Shirdi.

Annexure "A"

Requirement of 10 KVA UPS for Shri Saibaba Hospital.

(Preferred Make: Emerson, APC, Socomec, Libert or equivalent)

Sr. No.	Description	Qty.	Rate Per in Rs. (Incl. of all Taxes & Duties)	Total Amt. in Rs.
	<p>10 KVA UPS</p> <ul style="list-style-type: none">• IGBT based converter and Inverter• 100 % isolation between input and output• High AC to AC efficiency• Power factor >0.98• Battery : 12Volt/18 AH, SMF , ISI marked, Stand to be provided for battery• Battery charging should be possible even at low voltage up to 180 Volt• Back up: Minimum 1 Hr.• Input voltage: 170 Volt to 270 Volt, Single phase• Input Frequency:45 Hz to 60 Hz• Output : 230 V, 50Hz, Single Phase, Pure Sine wave, Stabilized with spike Guard Surge Protection• Output sockets: 230 V, 15 Amp. = 02 Nos.• Display with indicator for AC mode, Battery mode, Battery charging, low battery, Bypass, Overload, Fault, Input & output voltage, input & output frequency, % Load on UPS.• Output Voltage Regulation max. $\pm 1\%$• Output Frequency regulation : ± 0.1 Hz• Transfer Time: 0 msec.• Overload Capacity: 125% for 15 min. 150 % for 60 Seconds• Protection: Overload, Battery Overcharge, Battery Discharge, Output low, Short Circuit, Over Temperature, Surge• Noise Level: < 50 dB• Output must be provided through Inverter battery			

2010-11-T/Quotation-2010/4



**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India)
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi
Email : saibaba+hospital@saibai.org.in

Sr. No.	Description	Qty.	Rate Per in Rs. (Incl. of all Taxes & Duties)	Total Amt. in Rs.
	<ul style="list-style-type: none">All Technology should be mandatorily with safety approval - CE, IEC. Approval certificate is to be submitted in addition to CE Certificate for the total unit.The product should be time tested, branded and should be OEMWarranty-1YearsAMC/CMC Cost- AMC / CMC cost should be including Taxes. It should be fixed for next 5 years.Cost of Consumables to be fixed for warranty period.			
			Total Amt.in Rs.	

Medical Director,
Shri Saibaba Sansthan Trust,
Hospitals, Shirdi

